



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Loy Instrument, Inc.
8455 East 30th Street
Indianapolis, IN 46219

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to be 'Jason Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 17 March 2025

Certificate Number: L2079-1



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Loy Instrument, Inc.

8455 East 30th Street
Indianapolis, IN 46219
Stacey Atha
317-890-0474

CALIBRATION

Valid to: **March 17, 2025**

Certificate Number: **L2079-1**

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current	(4 to 20) mA	18 μ A	Digital Multimeter
DC Voltage	(0 to 100) mV (0 to 10) V	5.5 μ V 6.2 mV	
Electrical Simulation Thermocouple indicating devices	Type K (0 to 2 500) °F	0.64 °F	Precision Process Calibrator
	Type J (0 to 2 190) °F	0.64 °F	
	Type R (32 to 3 000) °F	0.66 °F	
	Type S (32 to 3 200) °F	0.7 °F	
	Type T (-320 to 750) °F	0.68 °F	
	Type N (0 to 2 370) °F	0.65 °F	
	Type B (500 to 3 000) °F	0.68 °F	
	Type C (600 to 4 200) °F	0.7 °F	
Electrical Simulation RTD indicating devices	PT100-385 (-200 to 850) °F	0.71 °F	Precision Process Calibrator


Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature Measure	32 °F	1.6 °F	Ice Bath and Type T Thermocouple AMS 2750G
Temperature System Accuracy Tests ¹	Types K, N (32 to 2 000) °F	2.5 °F	Reference Thermocouple with Readout unit AMS 2750G
Temperature Uniformity Surveys ¹	Types K, N (0 to 2 000) °F (2 000 to 2 400) °F	2.5 °F 4.2 °F	Reference Thermocouple with Digital Recorder AMS 2750G

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. L2079-1.



Jason Stine, Vice President

